

<b>Respirators <sup>A</sup></b>	<b>S1</b>	<b>S2</b>	<b>S3</b>	<b>P1</b>	<b>P2</b>	<b>P3</b>
3% Hydrogen Peroxide						
Average	0.59	0.36	ND	0.43	0.53	0.70
Lower 95% CI	0.14	0.28	—	0.12	0.20	0.38
Upper 95% CI	1.04	0.45	—	0.74	0.87	1.02
Vaporized Hydrogen Peroxide						
Average	1.23	0.43	0.36	1.09	0.81	0.35
Lower 95% CI	0.68	0.29	-0.11	0.64	0.29	0.04
Upper 95% CI	1.77	0.57	0.83	1.53	1.34	0.66
10% Bleach						
Average	0.37	0.70	ND	0.32	1.66	0.45
Lower 95% CI	0.00	0.29	—	-0.31	-2.03	-0.64
Upper 95% CI	0.73	1.11	—	0.95	5.34	1.54
Mixed Oxidants						
Average	0.14	0.08	ND	0.25	1.72	8.10
Lower 95% CI	-0.05	-0.08	—	-1.53	-1.38	3.06
Upper 95% CI	0.32	0.24	—	2.03	4.82	13.14
DMDO						
Average	7.38	7.72	4.53	5.53	7.19	5.14
Lower 95% CI	6.87	-0.09	2.52	5.11	6.50	3.95
Upper 95% CI	7.89	15.53	6.53	5.94	7.87	6.33

Note: ND = none detected (the detection limit (0.02 mL × 0.001 N reductant) is -0.02 µeq, -0.7 µg). <sup>A</sup> n = 3 for all samples in the table.